

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01724 Calib. Date: 09/21/2021 Calib. No.: 00055
Certification File No.: 01725 Cert. Date: 09/21/2021 Cert. No.: 00039
Linearity File No.: 01726 Lin. Date: 09/21/2021 Lin. No.: 00038
Solution File No.: 01737 Soln. Date: 11/15/2021 Soln. No.: 00295
Sequential File No.: 01737 File Date: 11/15/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363
Control Solution %: 0.100% Expires: 07/06/2022
Solution Control Lot: 20320 Bottle No.: 1463

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:53S		
Control 1 EC	0.098%	19:54S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:54S		
Control 2 EC	0.098%	19:55S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:56S		
Control 3 EC	0.099%	19:56S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:57S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: 

Badge No.: 68

Date: 11/15/2021