

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01694	Calib. Date: 03/31/2021
Certification File No.:	01695	Calib. No.: 00054
Linearity File No.:	01696	Cert. Date: 03/31/2021
Solution File No.:	01718	Cert. No.: 00038
Sequential File No.:	01718	Lin. Date: 03/31/2021
		Lin. No.: 00037
		Soln. Date: 08/16/2021
		Soln. No.: 00290
		File Date: 08/16/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	20320	Expires: 07/06/2022
		Bottle No.: 1469

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:45D		
Control 1 EC	0.097%	13:46D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47D		
Control 2 EC	0.097%	13:47D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48D		
Control 3 EC	0.097%	13:48D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	13:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: BROWN

First Name: KYLE

MI: W

Signature: \_\_\_\_\_



Badge No.: 089

Date: 08/16/2021