

# Calibrating Unit


## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01751	Calib. Date: 02/24/2022
Certification File No.:	01752	Cert. No.: 00056
Linearity File No.:	01753	Cert. Date: 02/24/2022
Solution File No.:	01764	Lin. No.: 00039
Sequential File No.:	01764	Lin. Date: 02/24/2022
		Soln. No.: 00304
		Soln. Date: 06/04/2022
		File Date: 06/04/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 1482

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:16D		
<del>Control 1 EC</del>	<del>0.099%</del>	<del>22:17D</del>	<del>33.9°C</del>	<del>*** TEST PASSED ***</del>
Control 1 IR	0.099%	22:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:18D		
Control 2 EC	0.098%	22:18D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:19D		
Control 3 EC	0.097%	22:19D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

<b>Changed By:</b>			
Last Name: MCQUADE		First Name: EDWARD	MI: J
Signature: _____	#97	Badge No.: 97	
		Date: 06/04/2022	

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01751	Calib. Date: 02/24/2022
Certification File No.:	01752	Calib. No.: 00056
Linearity File No.:	01753	Cert. Date: 02/24/2022
Solution File No.:	01763	Cert. No.: 00040
Sequential File No.:	01763	Lin. Date: 02/24/2022
		Lin. No.: 00039
		Soln. Date: 06/04/2022
		Soln. No.: 00303
		File Date: 06/04/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 1482

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:50D		
<del>Control 1 EC</del>	<del>-.---%</del>	<del>18:50D</del>	<del>24.0°C</del>	<del>SIMULATOR TEMP. ERROR</del>
Control 1 IR	-.---%	18:50D	24.0°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: \_\_\_\_\_



Badge No.: 97

Date: 06/04/2022

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01751	Calib. Date: 02/24/2022
Certification File No.:	01752	Calib. No.: 00056
Linearity File No.:	01753	Cert. Date: 02/24/2022
Solution File No.:	01762	Cert. No.: 00040
Sequential File No.:	01762	Lin. Date: 02/24/2022
		Lin. No.: 00039
		Soln. Date: 06/04/2022
		Soln. No.: 00302
		File Date: 06/04/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 1142

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:45D		
Control 1 EC	0.098%	13:46D	33.9°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	13:46D	33.9°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	13:47D		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: \_\_\_\_\_

Badge No.: 97

Date: 06/04/2022