

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01830	Calib. Date: 06/06/2023
Certification File No.:	01831	Calib. No.: 00059
Linearity File No.:	01832	Cert. Date: 06/06/2023
Solution File No.:	01833	Cert. No.: 00043
Sequential File No.:	01833	Lin. Date: 06/06/2023
		Lin. No.: 00042
		Soln. Date: 06/06/2023
		Soln. No.: 00319
		File Date: 06/06/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 1356

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:07D		
Control 1 EC	0.100%	15:08D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	15:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:09D		
Control 2 EC	0.099%	15:09D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:10D		
Control 3 EC	0.098%	15:10D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Temperature Probe Serial Number: DDWA P2-216 (8)

Changed By:

Last Name: GUARCELLO

First Name: COREY

MI: R.

Signature: 

Badge No.: 7035

Date: 06/06/2023