

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01830	Calib. Date:	06/06/2023
Certification File No.:	01831	Cert. Date:	06/06/2023
Linearity File No.:	01832	Lin. Date:	06/06/2023
Solution File No.:	01836	Soln. Date:	07/05/2023
Sequential File No.:	01836	File Date:	07/05/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	1489

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:00D		
Control 1 EC	0.100%	16:00D	33.9°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	16:00D	33.9°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	16:01D		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: _____

Edward J. McQuade #97

Badge No.: 97

Date: 07/05/2023