

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00953	Calib. Date: 09/18/2012
Certification File No.:	00954	Calib. No.: 00028
Linearity File No.:	00955	Cert. Date: 09/18/2012
Solution File No.:	00979	Cert. No.: 00016
Sequential File No.:	00979	Lin. Date: 09/18/2012
		Lin. No.: 00016
		Soln. Date: 01/10/2013
		Soln. No.: 00130
		File Date: 01/10/2013
Calibrating Unit:	1	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 0424

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:01S		
Control 1 EC	0.104%	12:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02S		
Control 2 EC	0.101%	12:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	12:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:03S		
Control 3 EC	0.101%	12:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:05S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRANCHAK

First Name: MARYLOUISE

MI: -

Signature: 

Badge No.: 431

Date: 01/10/2013