

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	00841	Calib. Date:	10/26/2011
Certification File No.:	00842	Cert. No.:	00026
Linearity File No.:	00843	Cert. Date:	10/26/2011
Solution File No.:	00872	Lin. Date:	10/26/2011
Sequential File No.:	00872	Lin. No.:	00014
		Soln. Date:	01/26/2012
		Soln. No.:	00114
		File Date:	01/26/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	10H081	Expires:	08/04/2012
		Bottle No.:	0186

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:24S		
Control 1 EC	0.100%	14:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:26S		
Control 2 EC	0.098%	14:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:27S		
Control 3 EC	0.098%	14:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Sgt. T. Tarr

Badge No.: 427

Date: 01/26/2012