

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01118	Calib. Date: 07/09/2014
Certification File No.:	01119	Calib. No.: 00032
Linearity File No.:	01120	Cert. Date: 07/09/2014
Solution File No.:	01148	Cert. No.: 00020
Sequential File No.:	01148	Lin. Date: 07/09/2014
		Lin. No.: 00020
		Soln. Date: 10/24/2014
		Soln. No.: 00163
		File Date: 10/24/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	13G115	Expires: 07/18/2015
		Bottle No.: 0010

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:15D		
Control 1 EC	0.099%	13:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:16D		
Control 2 EC	0.099%	13:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:18D		
Control 3 EC	0.099%	13:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature:  #585

Badge No.: 585

Date: 10/24/2014