

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01030	Calib. Date:	08/14/2013
Certification File No.:	01031	Cert. Date:	08/14/2013
Linearity File No.:	01032	Lin. Date:	08/14/2013
Solution File No.:	01044	Soln. Date:	10/03/2013
Sequential File No.:	01044	File Date:	10/03/2013
		Calib. No.:	00030
		Cert. No.:	00018
		Lin. No.:	00018
		Soln. No.:	00143
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	12L109	Expires:	12/26/2014
		Bottle No.:	0122

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:16D		
Control 1 EC	0.099%	11:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17D		
Control 2 EC	0.099%	11:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19D		
Control 3 EC	0.098%	11:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MCMAHON

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 80

Date: 10/03/2013