

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01030	Calib. Date:	08/14/2013
Certification File No.:	01031	Cert. No.:	00030
Linearity File No.:	01032	Lin. Date:	08/14/2013
Solution File No.:	01040	Cert. No.:	00018
Sequential File No.:	01040	Lin. No.:	00018
		Soln. Date:	09/06/2013
		Soln. No.:	00142
		File Date:	09/06/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	12L109	Expires:	12/26/2014
		Bottle No.:	0123

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:16D		
Control 1 EC	0.099%	17:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:17D		
Control 2 EC	0.099%	17:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:18D		
Control 3 EC	0.099%	17:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: \_\_\_\_\_



Badge No.: 68

Date: 09/06/2013