

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01278	Calib. Date: 12/09/2015
Certification File No.:	01279	Calib. No.: 00036
Linearity File No.:	01280	Cert. Date: 12/09/2015
Solution File No.:	01293	Cert. No.: 00023
Sequential File No.:	01293	Lin. Date: 12/09/2015
		Lin. No.: 00023
		Soln. Date: 01/22/2016
		Soln. No.: 00191
		File Date: 01/22/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	15D138	Expires: 04/13/2017
		Bottle No.: 0155

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:57S		
Control 1 EC	0.102%	11:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:58S		
Control 2 EC	0.101%	11:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59S		
Control 3 EC	0.101%	12:00S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.097%	12:00S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SABATINO

First Name: FRANK

MI: T

Signature: _____



Badge No.: 583

Date: 01/22/2016