

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01222	Calib. Date:	06/10/2015
Certification File No.:	01223	Cert. Date:	06/10/2015
Linearity File No.:	01224	Lin. Date:	06/10/2015
Solution File No.:	01246	Soln. Date:	08/29/2015
Sequential File No.:	01246	File Date:	08/29/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0364
Solution Control Lot:	15D138	Expires:	04/13/2017
		Bottle No.:	0164

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:22D		
Control 1 EC	0.100%	11:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23D		
Control 2 EC	0.099%	11:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25D		
Control 3 EC	0.099%	11:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: BROWN

First Name: KYLE

MI: W

Signature: \_\_\_\_\_



Badge No.: 089

Date: 08/29/2015