

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0382  
Location: WEST WINDSOR POLICE  
Calibration File No.: 01222 Calib. Date: 06/10/2015 Calib. No.: 00034  
Certification File No.: 01223 Cert. Date: 06/10/2015 Cert. No.: 00022  
Linearity File No.: 01224 Lin. Date: 06/10/2015 Lin. No.: 00022  
Solution File No.: 01263 Soln. Date: 10/20/2015 Soln. No.: 00184  
Sequential File No.: 01263 File Date: 10/20/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364  
Control Solution %: 0.100% Expires: 04/13/2017  
Solution Control Lot: 15D138 Bottle No.: 0174

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:26D		
Control 1 EC	0.100%	09:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:27D		
Control 2 EC	0.099%	09:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:29D		
Control 3 EC	0.099%	09:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


## Changed By:

Last Name: TARR

First Name: MARYLOUISE

MI: -

Signature: \_\_\_\_\_



Badge No.: 431

Date: 10/20/2015