

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01513 Calib. Date: 04/27/2018 Calib. No.: 00043
Certification File No.: 01514 Cert. Date: 04/27/2018 Cert. No.: 00030
Linearity File No.: 01515 Lin. Date: 04/27/2018 Lin. No.: 00029
Solution File No.: 01530 Soln. Date: 06/16/2018 Soln. No.: 00233
Sequential File No.: 01530 File Date: 06/16/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363
Control Solution %: 0.100% Expires: 09/14/2019
Solution Control Lot: 17290 Bottle No.: 0633

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:54D		
Control 1 EC	0.099%	08:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55D		
Control 2 EC	0.099%	08:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:56D		
Control 3 EC	0.100%	08:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____



Badge No.: 568

Date: 06/16/2018