

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01579	Calib. Date: 03/07/2019
Certification File No.:	01580	Calib. No.: 00045
Linearity File No.:	01581	Cert. Date: 03/07/2019
Solution File No.:	01602	Cert. No.: 00032
Sequential File No.:	01602	Lin. Date: 03/07/2019
		Lin. No.: 00031
		Soln. Date: 07/02/2019
		Soln. No.: 00252
		File Date: 07/02/2019

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	18320	Expires: 09/26/2020
		Bottle No.: 0784

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:31D		
Control 1 EC	0.101%	11:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32D		
Control 2 EC	0.100%	11:33D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34D		
Control 3 EC	0.100%	11:34D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____



Badge No.: 68

Date: 07/02/2019